

## 新北市政府衛生局 函

地址：220205新北市板橋區英士路192之1號3樓  
承辦人：江佳穎  
電話：(02)22577155 分機2353  
傳真：(02)22536548  
電子信箱：AQ5750@ntpc.gov.tw



24158

新北市三重區重新路5段646號8樓

受文者：新北市藥師公會

發文日期：中華民國110年8月2日  
發文字號：新北衛食字第1101382392號  
速別：普通件  
密等及解密條件或保密期限：  
附件：產品照片1份

主旨：有關澄希貿易有限公司輸入銷售之「工廠直貨直銷LK87數碼館飽和度指脈氧手指夾式脈搏監測儀血氧機(一般型)」及「FDA/CE指夾式OLED屏血氧飽和度監測儀指尖脈搏心律血氧儀」產品回收一案，請轉知所屬會員，如有案內產品請立即下架勿再販售與使用，並儘速配合回收事宜，請查照。

說明：

- 一、依據桃園市政府衛生局110年7月21日桃衛藥字第1100062132號函辦理。
- 二、案係臺灣士林地方檢察署、法務部調查局臺北市調查處、衛生福利部食品藥物管理署及臺北市政府衛生局至旨揭公司辦公場所查獲旨揭產品，該產品皆未經查驗登記取得醫療器材許可證，涉違反醫療器材管理法第62條規定，故啟動回收。
- 三、本案係屬第一級回收，為保障民眾用藥安全，請轉知所屬會員，如有案內批號產品請立即下架勿再販售與使用，並儘速配合回收事宜。

正本：社團法人新北市醫師公會、新北市藥師公會、新北市藥劑生公會、新北市醫療器材商業同業公會、新北市西藥商業同業公會、新北市商業會

副本：

局長 陳潤秋

本案依分層負責規定授權業務主管決行

FOR HEALTH FOR YOU

RoHS CE FCC E.A.

FINGERTIP PULSE  
OXIMETER



117-87

**AUTOMATIC SHUTDOWN  
FOR MORE ENERGY SAVING**

Automatic reactivation after finger position  
adjustment. Automatic shutdown after 8  
seconds without finger.  
Energy saving (long term standby).

Original Quality

**AUTOMATIC SHUTDOWN  
FOR MORE ENERGY SAVING**

Automatic reactivation of finger position,  
automatic shutdown after 8  
seconds without finger.  
Energy saving (long term standby).

Original Quality

